

COMMON SET OF DEVIATIONS

Tender Notice No.: *GSIDC/ ENGG./NIT-103/2019-2020 dated 17th December 2019 - Line A1*Pre-bid queries last date of submission: *23rd December 2019*

Sr. No.	Volume / Chapter/ Clause / Item No. / Page No.	Clarification / changes sought by Prospective Bidders	Decision of GSIDC
1.	Volume II, Item No. A1, (bullet pt.4), Page 7	Joints between the panels in the Operation Theatre should be sealed by using tongue and groove technology.	Strictly as per tender
2.	Volume II, Item No. A1, (bullet pt.4), Page 7 & 8	The gaskets to be made of non- toxic silicon in compliance with regulations applicable to clean rooms, providing a durable and non-degradable seal. This is to be deleted.	Strictly as per tender
3.	Volume II, Item No. A1, (bullet pt.8), Page 8	Wall to Wall & Wall to Ceiling coving should be made of stainless steel 316 grade/ Aluminium	Strictly as per tender
4.	Volume II, Item No. A1, (bullet pt.10), Page 8	The panel joints shall be sealed with silicon sealant / epoxy metal filler	Acceptable
5.	Volume II, Item No. A1, (bullet pt.12), Page 8	Ceiling should be 1mm thick SS 316 grade in the front side and 0.8mm thick GI sheet on the back side of panel with PUF insulated prefabricated SS panels. Thickness of ceiling panel should be 50mm thick.	Ceiling should be min 1mm thick SS 316 grade in the front side and 0.8mm thick GI sheet on the back side of panel with PUF insulated prefabricated SS panels. Thickness of ceiling panel should be minimum 40mm thick.

COMMON SET OF DEVIATIONS

Sr. No.	Volume / Chapter/ Clause / Item No. / Page No.	Clarification / changes sought by Prospective Bidders	Decision of GSIDC
6.	Volume II, Item No. A1, Page 9	<u>Approved Makes:</u> Ceiling/ Wall panel – Volantis / PES / Mediconz International / MDD India	Acceptable
7.	Volume II, Item No. A2, Page 9	<u>Approved Makes:</u> Volantis / PES / Mediconz International / MDD India	Acceptable
8.	Volume II, Item No. A4-i, (pt. 4), Page 12	Door panels shall be 58-60 mm thick, 0.6mm sandwiched panel on both side & 58.5mm PU sandwiched between the panels of SS 304	Not Acceptable. Strictly as per tender
9.	Volume II, Item No. A4-ii, (Pt. 1& 3), Page 14	The size of Main OT Door shall be 1500 X 2100 mm finished with Antimicrobial High Pressure Laminate on both sides	Acceptable
10.	Volume II, Item No. A4-ii, (Pt. 4), Page 14	HPL finish in 6 colours	External finish shall be of D finish / hair line finish/ HPL finish in 6 colours.
11.	Volume II, Item No. A4-ii, (Pt. 5), Page 14	Vision panels of size 300mm x 300mm with 44 mm double glare thick glass.	Vision panels of size 300mm x 300mm with 4-6 mm double glare thick glass.
12.	Volume II, Item No. A4-ii, (Pt. 8), Page 14	Should have door closer used on one door panel and with door handles and hinges.	Strictly as per tender

COMMON SET OF DEVIATIONS

Sr. No.	Volume / Chapter/ Clause / Item No. / Page No.	Clarification / changes sought by Prospective Bidders	Decision of GSIDC
13.	Volume II, Item No. A4-ii, (Pt. 9), Page 14	Should have one pair of D handle, 1 door stopper with hold open arm, 1 set of lock with 3 keys for each set of doors.	Acceptable
14.	Volume II, Item No. A4-ii, Page 14	<u>Approved Makes:</u> Alvo / Infimed / Volantis / PES / Metaflex / Ditec Entrematic / Mediconz International / SHD Italy/ Kone / MDD India	Acceptable
15.	Volume II, Item No. A5 (Pt. 7), Page 15	Overall size of the P.R.D: 325 x 345 mm approx	Overall size of the P.R.D shall be (305-325mm) x (335-345mm)
16.	Volume II, Item No. A5, Page 15	<u>Approved Makes:</u> Volantis / Mediconz International / PES / MDD India	Acceptable
17.	Volume II, Item No. A6, Page 15	To mention the size of peripheral light as : 2’ x 1’ approx	Size of Peripheral lights is specified in Volume III. Strictly as per tender
18.	Volume II, Item No. A6, Page 15	<u>Approved Makes:</u> Philips / Wipro / Volantis / Mediconz International / MDD India	Acceptable
19.	Volume II, Item No. A7 (last line), Page 15	It shall be made of 1.2mm GI powder coated body with front frame in Extruded Aluminum SS 304 G, 1.2mm sheet	Strictly as per tender

COMMON SET OF DEVIATIONS

Sr. No.	Volume / Chapter/ Clause / Item No. / Page No.	Clarification / changes sought by Prospective Bidders	Decision of GSIDC
20.	Volume II, Item No. A7, Page 15	<u>Approved Makes:</u> PES / Mediconz International/ Volantis / MDD India or equivalent	Acceptable
21.	Volume II, Item No. A8, (pt.2), Page 16	Manufacturers Test Certificate to be provided along with supply of material instead of Government approved third party certification.	Not Acceptable. Strictly as per tender
22.	Volume II, Item No. A9, Page 17	<u>Approved Makes:</u> PES / Mediconz International/ Volantis / MDD India or equivalent	Acceptable
23.	Volume II, Item No. A11, Page 18	It should be 21” touch screen type control panel	Minimum 19” LCD – TFT type Touch screen control panel
24.	Volume II, Item No. A11, (bullet pt. 1), Page 18	Digital clock	Strictly as per tender
25.	Volume II, Item No. A11, (bullet pt. 13, 15 &17), Page 19	1 Touch Switch for X-ray Viewer 1 Open Touch Switch for additional device control Power Failure Message These features to be deleted	Not Acceptable. Strictly as per tender

COMMON SET OF DEVIATIONS

Sr. No.	Volume / Chapter/ Clause / Item No. / Page No.	Clarification / changes sought by Prospective Bidders	Decision of GSIDC
26.	Volume II, Item No. A11, Page 19	<u>Approved Makes:</u> Alvo / Infimed / PES / Trumf / Stryker / Mediconz International / Volantis/ MDD India	Acceptable
27.	Volume II, Item No. A12, (pt. 3), Page 20	It should have colour temperature with adjustments of (+/- 10%) of 3700 -5000K to cover maximum colour temperatures of different companies	Color temperature Tc (with adjustment) – 3500 - 5000K
28.	Volume II, Item No. A12, (pt. 6), Page 20	CRI should be > 90 > 95	Color Rendering Index [Ra(1-8)]->90(>95)
29.	Volume II, Item No. A12, Page 20	<u>Approved Makes:</u> Alvo / Infimed / PES / Drager / Volantis / Maquet / Dr Mach / Steris / MDD India	Acceptable
30.	Volume II, Item No. A13, Page 20	<u>Approved Makes:</u> Alvo / Infimed / PES / Drager / Volantis / Maquet / Dr Mach / Steris / MDD India	Acceptable
31.	Volume II, Item No. A14, Page 21	<u>Approved Makes:</u> Volantis / Mediconz International / PES / MDD India or equivalent	Acceptable

COMMON SET OF DEVIATIONS

Sr. No.	Volume / Chapter/ Clause / Item No. / Page No.	Clarification / changes sought by Prospective Bidders	Decision of GSIDC
32.	Volume II, Item No. A15, Page 21	<u>Approved Makes:</u> Volantis / Mediconz International / PES / MDD India or equivalent	Acceptable
33.	Volume II, Item No. A16(a), Page 21	<u>Approved Makes:</u> Volantis / Mediconz International / PES/ MDD India	Acceptable
34.	Volume II, Item No. A16(b), Page 23	<u>Approved Makes:</u> Volantis / Mediconz International / PES/ MDD India	Acceptable
35.	Volume II, Item No. A17, Page 24	<u>Approved Makes:</u> Volantis / Mediconz International / PES/ MDD India	Acceptable
36.	Volume II, Manufacturer Quality Standards & Certification, (pt. 3), Page 24	Should be amended to ISO or CE certification	Strictly as per tender
37.	Volume II, Manufacturer Quality Standards & Certification, (pt. 3), Page 24	The manufacturer/ bidder should compliant with ISO 9001, ISO 13485, ISO 14001, OHSAS 18001.	Strictly as per tender